



DET NORSKE VERITAS  
CERTIFICATION AS

*Compliance Testing & Field  
Measurement*

Veritasveien 1  
N-1322 HØVIK Norway  
Tel: +47 67 57 99 00  
Fax: +47 67 57 99 11  
<http://www.dnv.com>

RAE Systems Europe  
Kirstinehøj 23A

2770 Kastrup  
Danmark

Att: Ole Olsen

Your ref.:  
Ole Olsen

Our ref.:  
dinh

Date:  
2010-12-14

### Statement of testing

**IP65 testing of ToxiRAE Pro Gas Detectors, version ToxiRAE Pro PGM-1860 according to IEC 60529 ED 2.1 (2001).**

Hereby confirm that the following tests were performed at the Environmental testing laboratory of Det Norske Veritas Certification AS on the time period from 2010-11-19 to 2010-10-19:

- 1) Chapter 13.5 – Special conditions for first characteristic numeral 6.
- 2) Chapter 14.2.4 – Test for second characteristic numeral 5 with spray nozzle.

The detectors passed the tests, see DNV test report 2010-3608.

This statement also covers version ToxiRAE Pro PGM-1820 with same mechanical construction.

Yours faithfully  
for DET NORSKE VERITAS CERTIFICATION AS

Einar Richter Jordfald  
Head of section  
Compliance Testing & Field Measurement

Huu Dung Dinh  
Saksbehandler